



MM Scale Co., Ltd.

บริษัท เอ็มเอ็ม สเกล จำกัด

1 ซอยเทียนทะเล 20 แขวง 8-5-2 ถนนบางขุนเทียน-ชายทะเล แขวงแสมดำ เขตบางขุนเทียน กรุงเทพมหานคร 10150
Tel. 094-671-8723

CALIBRATION REPORT

Calibration Report No ร้าง — ออกเลขเมื่ออนุมัติ

Customer : ทดสอบ ทดสอบ (สำนักงานใหญ่)

Address :

Tel : Fax :

Tax identification number :

Manufacturer : ทดสอบ

ID No : FA-202605213

Model : ASDG651561

SERIAL NO : 551238561ASD6358A
Sd

Max Capacity : 6000 g.

D : 0.1 g.

Accuracy Class : 97%

Standards Used : 20 kg

Serviced and Tested By :

Repeatability of Reading :

Reading	As Found of Reading	After Adjustment of Readings
6000 g.	5998.0 g.	6000.0 g.
6000 g.	5998.0 g.	6000.0 g.

Linearity of Reading :

shift Test	As Found	After Adjustment
600 g.	600.0 g.	600.0 g.
1200 g.	1200.0 g.	1200.0 g.
1800 g.	1800.0 g.	1800.0 g.
2400 g.	2400.0 g.	2400.0 g.
3000 g.	3000.0 g.	3000.0 g.
3600 g.	3600.0 g.	3600.0 g.
4200 g.	4200.0 g.	4200.0 g.
4800 g.	4800.0 g.	4800.0 g.
5400 g.	5400.0 g.	5400.0 g.
6000 g.	6000.0 g.	6000.0 g.

off-Center Loading :

Center	Front	Back	Left	Right	Uncer Maximum point
6000.0 g.	6000.0 g.	6000.0 g.	6000.0 g.	6000.0 g.	± 0.1 g.

Corner Loading :

Top-Left	Top-Right	Bottom-Left	Bottom-Right
6000.0 g.	6000.0 g.	6000.0 g.	6000.0 g.

This calibration has been done with approved weights according to W&M. Testing procedure according to standard testing procedure of MM Scale.

This report does not extend the scale's calibration period.

Authorized Signature

Cal. Date 12/10/2569

Calibration due (per last accredited calibration) —

Tested and Issued By :
MM Scale Co., Ltd.



MM-00001 - app.mm-scale.com/t/MM-00001